

Listing #: E212383  
Report #: 27861  
Original Certification Date: January 25, 2010  
Revised Certification Date: October 21, 2025

This Certification is issued to:  
Advanced Micro Instruments, Inc.  
15501 Red Hill Ave Ste 100  
Tustin, CA 92780

Stating that the product(s):  
Oxygen Deficiency Monitor  
Model 221R

Product Rating(s):  
• 90-230VAC, 60Hz, 15W

Achieved Certification to the following standard(s):  
UL 61010-1 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements, Edition: 3; Edition Date: May 11, 2012; Rev. July 19, 2019



Hon Keung Ip  
Certification Officer,  
Eurofins Electrical and Electronic Testing North America, LLC



*All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized by Eurofins E&E NA to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.*

*Eurofins E&E North America, LLC is recognized by OSHA and accredited by the Standards Council of Canada.*

**N**  
**R**  
**T**  
**L**